

耀創工程有限公司

Sun Creation Engineering Limited **Calibration & Testing Laboratory**

Certificate of Calibration 校正證書

Certificate No.:

C193125

證書編號

ITEM TESTED / 送檢項目 (Job No. / 序引編號: IC19-1166)

Date of Receipt / 收件日期: 10 June 2019

Description / 儀器名稱

Sound Level Meter

Manufacturer / 製造商 Model No. / 型號

Castle

Serial No. / 編號

GA116I 070784

Supplied By / 委託者

Sam & Sam Testing & Surveyor Services Co., Ltd.

Room 40, 8/F., Sino Industrial Plaza,

9 Kai Cheung Road, Kowloon Bay, Kowloon

TEST CONDITIONS / 測試條件

Temperature / 温度 :

 $(23 \pm 2)^{\circ}C$

Relative Humidity / 相對濕度 :

 $(50 \pm 25)\%$

Line Voltage / 電壓

TEST SPECIFICATIONS / 測試規範

Calibration check

DATE OF TEST / 測試日期

17 June 2019

TEST RESULTS / 測試結果

The results apply to the particular unit-under-test only.

The results do not exceed manufacturer's specification.

The results are detailed in the subsequent page(s).

The test equipment used for calibration are traceable to National Standards via:

- The Government of The Hong Kong Special Administrative Region Standard & Calibration Laboratory
- The Bruel & Kjaer Calibration Laboratory, Denmark
- Agilent Technologies / Keysight Technologies
- Fluke Everett Service Center, USA

Tested By 測試

K P Cheuk Assistant Engineer

Certified By 核證

K C/Lee Engineer

Date of Issue 簽發日期

19 June 2019

The test equipment used for calibration are traceable to the Nation Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory,

本證書所載校正用之測試器材均可測源至國際標準。局部複印本證書需先獲本實驗所書面批准。



輝創工程有限公司

Sun Creation Engineering Limited **Calibration & Testing Laboratory**

Certificate of Calibration 校正證書

Certificate No.: C193125

證書編號

- 1. The unit-under-test (UUT) was allowed to stabilize in the laboratory for over 12 hours, and switched on to warm up for over 10 minutes before the commencement of the test.
- 2. Self-calibration using the laboratory acoustic calibrator was performed before the test from 6.1.1.2 to 6.3.2.
- 3. The results presented are the mean of 3 measurements at each calibration point.
- 4. Test equipment:

Equipment ID

Description

Certificate No.

CL280 CL281

40 MHz Arbitrary Waveform Generator

C190176

Multifunction Acoustic Calibrator

CDK1806821

- 5. Test procedure: MA101N.
- 6. Results:
- 6.1 Sound Pressure Level
- 6.1.1 Reference Sound Pressure Level

6.1.1.1 Before Self-calibration

		UUT	Setting	Applied	UUT		
Í	Range Parameter Frequency Time				Level	Freq.	Reading
	(dB)		Weighting	Weighting	(dB)	(kHz)	(dB)
- 1	55 - 120	L	Α	F	94.00	1	93.8

6.1.1.2 After Self-calibration

	UUT	Setting	Applie	d Value	UUT	IEC 61672	
Range	Parameter	Frequency	Time	Level	Freq.	Reading	Class 1 Spec.
(dB)		Weighting	Weighting	(dB)	(kHz)	(dB)	(dB)
55 - 120	L	A	F	94.00	1	94.0	± 1.1

6.1.2 Linearity

	UUT	Setting	Applied	UUT		
			Time Weighting	Level (dB)	Freq. (kHz)	Reading (dB)
55 - 120	L	A	F	94.00 104.00	1	94.0 (Ref.) 103.9
				114.00		113.9

Website/網址: www.suncreation.com

IEC 61672 Class 1 Spec. : \pm 0.6 dB per 10 dB step and \pm 1.1 dB for overall different.

The test equipment used for calibration are traceable to the Nation Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗所書面批准。



輝創工程有限公司

Sun Creation Engineering Limited **Calibration & Testing Laboratory**

Certificate of Calibration

校正證書

Certificate No.: C193125

證書編號

6.2 Time Weighting

	UUT	Setting		Applied Value		UUT	IEC 61672
Range (dB)	~ · · · ·		Time Weighting	Level Freq. (dB) (kHz)		Reading (dB)	Class 1 Spec. (dB)
55 - 120	L	A	F	94.00	1	94.0	Ref.
			S			94.0	± 0.3

6.3 Frequency Weighting

6.3.1 A-Weighting

UUT Setting					lied Value	UUT	IEC 61672
Range	Parameter	Frequency	Time	Level	Freq.	Reading	Class 1 Spec.
(dB)		Weighting	Weighting	(dB)		(dB)	(dB)
55 - 120	L	Α	F	94.00	63 Hz	67.9	-26.2 ± 1.5
					125 Hz	77.8	-16.1 ± 1.5
					250 Hz	85.3	-8.6 ± 1.4
					500 Hz	90.7	-3.2 ± 1.4
					1 kHz	94.0	Ref.
					2 kHz	95,2	$+1.2 \pm 1.6$
]		4 kHz	95.0	+1.0 ± 1.6
					8 kHz	92.8	-1.1 (+2.1; -3.1)
					12.5 kHz	89.7	-4.3 (+3.0; -6.0)

6.3.2 C-Weighting

	UUT Setting				lied Value	UUT	IEC 61672
Range	Parameter	Frequency	Time	Level	Freq.	Reading	Class 1 Spec.
(dB)		Weighting	Weighting	(dB)		(dB)	(dB)
55 - 120	L	С	F	94.00	63 Hz	93.5	-0.8 ± 1.5
					125 Hz	93.9	-0.2 ± 1.5
					250 Hz	94.0	0.0 ± 1.4
					500 Hz	94.0	0.0 ± 1.4
					1 kHz	94.0	Ref.
					2 kHz	93.8	-0.2 ± 1.6
					4 kHz	93.2	-0.8 ± 1.6
ľ					8 kHz	90.9	-3.0 (+2.1; -3.1)
					12.5 kHz	87.7	-6.2 (+3.0 ; -6.0)

The test equipment used for calibration are traceable to the Nation Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory

本證書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗所書面批准。



輝創工程有限公司

Sun Creation Engineering Limited **Calibration & Testing Laboratory**

Certificate of Calibration 校正證書

Certificate No.: C193125

證書編號

Remarks: - UUT Microphone Model No.: ACO 7146A & S/N: 66027

- Mfr's Spec. : IEC 61672 Class 1

- Uncertainties of Applied Value: 94 dB : 63 Hz - 125 Hz $: \pm 0.35 \text{ dB}$

> 250 Hz - 500 Hz $: \pm 0.30 \text{ dB}$ 1 kHz $; \pm 0.20 \text{ dB}$ 2 kHz - 4 kHz $: \pm 0.35 \, dB$ 8 kHz $: \pm 0.45 \text{ dB}$ 12.5 kHz $: \pm 0.70 \text{ dB}$

104 dB : 1 kHz $: \pm 0.10 \text{ dB (Ref. 94 dB)}$ 114 dB : 1 kHz $: \pm 0.10 \text{ dB (Ref. 94 dB)}$

Note:

Only the original copy or the laboratory's certified true copy is valid.

The values given in this Certificate only relate to the values measured at the time of the test and any uncertainties quoted will not include allowance for the equipment long term drift, variations with environment changes, vibration and shock during transportation, overloading, mis-handling, or the capability of any other laboratory to repeat the measurement. Sun Creation Engineering Limited shall not be liable for any loss or damage resulting from the use of the equipment.

The test equipment used for calibration are traceable to the Nation Standards as specified in this certificate. This certificate shall not be reproduced except in full, without the prior written approval of this laboratory

本驚書所載校正用之測試器材均可溯源至國際標準。局部複印本證書需先獲本實驗所書面批准。

⁻ The uncertainties are for a confidence probability of not less than 95 %.